

# ABX Minotrol Retic



**LOT** RX 033  
Rev 1

**CONTROL**



**2013-05-05**  
(YYYY - MM - DD)

		<b>ABX</b>											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual													
RETIC	%	1.3	1.3	1.3	± 0.7	5.7	5.7	5.7	± 2.3	10.7	10.7	10.7	± 3.2
RETIC	%	1.63	1.52	1.52	± 0.82	4.92	5.08	5.08	± 1.27	10.10	10.40	10.40	± 2.10
RETIC	10 <sup>6</sup> /µl	0.056	0.053	0.053	± 0.028	0.168	0.174	0.174	± 0.044	0.344	0.355	0.355	± 0.071
MFI MRI *		28.6	30.6	30.6	± 8.0	43.9	43.6	43.6	± 10.0	46.4	47.5	47.5	± 10.0
GR RBC	10 <sup>6</sup> /µl	3.42	3.46	3.46	± 0.30	3.42	3.42	3.42	± 0.30	3.41	3.41	3.41	± 0.30
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5

9930075-A  
Ref: TEMP-0382 Rev.30

\* For research use only in Canada and USA