

## ABX Minotrol Retic



**LOT**

**RX 110**

**CONTROL**

(Exp.) **2011-01-05**  
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA			PENTRA	PENTRA			PENTRA	PENTRA		
		120 RETIC	DX120	120 RETIC	DX120	120 RETIC	DX120						
Méthode Manuelle / Manual													
RETIC	%	<b>1.3</b>	<b>1.3</b>		± 0.7	<b>5.8</b>	<b>5.8</b>		± 2.3	<b>10.9</b>	<b>10.9</b>		± 3.3
RETIC	%	<b>1.54</b>	<b>1.70</b>		± 0.85	<b>5.18</b>	<b>5.46</b>		± 1.37	<b>10.60</b>	<b>10.90</b>		± 2.20
RETIC	10 <sup>9</sup> /µl	<b>0.051</b>	<b>0.057</b>		± 0.029	<b>0.172</b>	<b>0.184</b>		± 0.046	<b>0.345</b>	<b>0.362</b>		± 0.072
MFI MRI *		<b>29.1</b>	<b>28.9</b>		± 8.0	<b>38.8</b>	<b>38.7</b>		± 10.0	<b>41.4</b>	<b>41.2</b>		± 10.0
GR RBC	10 <sup>9</sup> /µl	<b>3.34</b>	<b>3.38</b>		± 0.30	<b>3.32</b>	<b>3.37</b>		± 0.30	<b>3.25</b>	<b>3.32</b>		± 0.30
PIC RANGE		<b>25.5</b>	<b>25.5</b>		± 4.5	<b>25.5</b>	<b>25.5</b>		± 4.5	<b>25.5</b>	<b>25.5</b>		± 4.5

\* For research use only in Canada and USA

9930034-C  
Ref: FORM-0382 Rev.06