

ABX Minotrol Retic



LOT

RX 011

CONTROL

(Exp.) **2011-03-05**
 (YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA			PENTRA	PENTRA			PENTRA	PENTRA		
120 RETIC	DX120		120 RETIC	DX120		120 RETIC	DX120						
Méthode Manuelle / Manual RETIC	%	1.3	1.3		± 0.7	6.1	6.1		± 2.4	11.1	11.1		± 3.3
RETIC	%	1.20	1.27		± 0.64	5.02	5.08		± 1.27	10.80	11.00		± 2.20
RETIC	10 ⁶ /µl	0.041	0.044		± 0.022	0.170	0.172		± 0.043	0.372	0.381		± 0.076
MFI MRI *		22.3	22.1		± 8.0	38.1	38.6		± 10.0	42.9	43.1		± 10.0
GR RBC	10 ⁶ /µl	3.43	3.44		± 0.30	3.38	3.39		± 0.30	3.44	3.46		± 0.30
PIC RANGE		25.5	25.5		± 4.5	25.5	25.5		± 4.5	25.5	25.5		± 4.5

* For research use only in Canada and USA

9930034-C

Ref: FORM-0382 Rev.06