


LOT

RX 031

CONTROL

 (Exp.) **2011-05-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
120 RETIC	DX120		120 RETIC	DX120		120 RETIC	DX120						
Méthode Manuelle / Manual RETIC	%	1.3	1.3		± 0.7	6.1	6.1		± 2.4	11.0	11.0		± 3.3
RETIC	%	1.03	1.16		± 0.58	5.48	5.67		± 1.42	10.70	10.90		± 2.20
RETIC	10 ⁹ /µl	0.035	0.040		± 0.020	0.189	0.195		± 0.049	0.371	0.382		± 0.076
MFI MRI *		24.7	25.8		± 8.0	44.1	46.2		± 10.0	46.1	48.8		± 10.0
GR RBC	10 ⁶ /µl	3.41	3.42		± 0.30	3.45	3.44		± 0.30	3.47	3.50		± 0.30
PIC RANGE		25.5	25.5		± 4.5	25.5	25.5		± 4.5	25.5	25.5		± 4.5

9930034-C

Ref: FORM-0382 Rev.06

* For research use only in Canada and USA