

ABX Minotrol Retic



LOT

RX 092

CONTROL



(Exp.)

2012-11-05

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1	2			3							
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX 120	DX Nexus	120 RETIC	DX 120	DX Nexus	120 RETIC	DX 120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	5.6	5.6	5.6	± 2.2	10.5	10.5	10.5	± 3.2
RETIC	%	1.36	1.52	1.52	± 0.76	4.78	4.90	4.90	± 1.23	10.00	10.00	10.00	± 2.00
RETIC	10 ⁹ /µl	0.047	0.052	0.052	± 0.026	0.169	0.173	0.173	± 0.043	0.363	0.361	0.361	± 0.073
MFI MRI *		28.5	28.4	28.4	± 8.0	38.3	38.9	38.9	± 10.0	41.6	42.0	42.0	± 10.0
GR RBC	10 ⁹ /µl	3.42	3.41	3.41	± 0.30	3.54	3.53	3.53	± 0.30	3.63	3.61	3.61	± 0.30
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5

9930059-A

Ref: TEMP-0382 Rev.09

* For research use only in Canada and USA