

# ABX Minotrol Retic



**LOT** RX 013  
Rev 1

**CONTROL**

(Exp.) **2013-03-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1	2			3							
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual													
RETIC	%	1.6	1.6	1.6	± 0.8	5.5	5.5	5.5	± 2.2	11.0	11.0	11.0	± 3.3
RETIC	%	2.13	2.08	2.08	± 1.07	5.26	5.29	5.29	± 1.32	10.50	10.90	10.90	± 2.20
RETIC	10 <sup>9</sup> /µl	0.075	0.072	0.072	± 0.038	0.186	0.183	0.183	± 0.047	0.369	0.378	0.378	± 0.076
MFI MRI *		35.6	35.3	35.3	± 8.0	42.3	42.2	42.2	± 10.0	44.7	45.2	45.2	± 10.0
GR RBC	10 <sup>9</sup> /µl	3.53	3.48	3.48	± 0.30	3.54	3.46	3.46	± 0.30	3.51	3.47	3.47	± 0.30
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5

9930059-A  
Ref: TEMP-0382 Rev.10

\* For research use only in Canada and USA