

# ABX Minotrol Retic



**LOT** RX 073

**CONTROL**

(Exp.) 2013-09-05  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1	2			3							
		PENTRA 120 RETIC	PENTRA DX120	PENTRA DX Nexus		PENTRA 120 RETIC	PENTRA DX120	PENTRA DX Nexus		PENTRA 120 RETIC	PENTRA DX120	PENTRA DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	6.2	6.2	6.2	± 2.5	10.9	10.9	10.9	± 3.3
RETIC	%	2.06	1.85	1.85	± 1.03	4.94	4.96	4.96	± 1.24	10.10	10.00	10.00	± 2.00
RETIC	10 <sup>6</sup> /µl	0.071	0.065	0.065	± 0.036	0.168	0.172	0.172	± 0.043	0.342	0.344	0.344	± 0.069
MFI MRI *		27.5	29.9	29.9	± 8.0	36.8	38.1	38.1	± 10.0	41.8	40.5	40.5	± 10.0
GR RBC	10 <sup>6</sup> /µl	3.45	3.49	3.49	± 0.30	3.41	3.46	3.46	± 0.30	3.39	3.44	3.44	± 0.30
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.31 FRONT / RECTO 9930083 -A

# ABX Minotrol Retic



**LOT** RX 073

**CONTROL**

 (Exp.) 2013-09-05  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX												
		CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		
		XLR				XLR				XLR				
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>6.2</b>			± 2.5	<b>10.9</b>			± 3.3	
RETIC	%	<b>1.89</b>			± 0.95	<b>5.13</b>			± 1.28	<b>10.30</b>			± 2.10	
RETIC	10 <sup>6</sup> /µl	<b>0.065</b>			± 0.033	<b>0.173</b>			± 0.043	<b>0.345</b>			± 0.069	
MFI MRI *		<b>27.6</b>			± 8.0	<b>38.2</b>			± 10.0	<b>41.6</b>			± 10.0	
GR RBC	10 <sup>6</sup> /µl	<b>3.43</b>			± 0.30	<b>3.38</b>			± 0.30	<b>3.35</b>			± 0.30	
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4	

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.31 BACK / VERSO 9930083 -A