

ABX Minotrol Retic



LOT RX 014
Rev 1

CONTROL

 (Exp.) 2014-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual													
RETIC	%	1.5	1.5	1.5	± 0.8	6.1	6.1	6.1	± 2.4	10.9	10.9	10.9	± 3.3
RETIC	%	1.90	1.85	1.85	± 0.95	5.24	5.17	5.17	± 1.31	10.90	11.00	11.00	± 2.20
RETIC	10 ⁶ /µl	0.065	0.064	0.064	± 0.033	0.182	0.180	0.180	± 0.046	0.373	0.378	0.378	± 0.076
MFI MRI *		32.0	31.2	31.2	± 8.0	43.7	41.9	41.9	± 10.0	45.6	45.0	45.0	± 10.0
GR RBC	10 ⁶ /µl	3.44	3.46	3.46	± 0.30	3.47	3.48	3.48	± 0.30	3.42	3.44	3.44	± 0.30
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.31 FRONT / RECTO 9930083 -A

ABX Minotrol Retic



LOT RX 014

CONTROL

(Exp.) 2014-03-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	6.1			± 2.4	10.9			± 3.3
RETIC	%	1.87			± 0.95	4.98			± 1.31	10.50			± 2.20
RETIC	10 ⁶ /µl	0.064			± 0.033	0.173			± 0.046	0.360			± 0.076
MFI MRI *		32.6			± 8.0	43.3			± 10.0	45.8			± 10.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.47			± 0.30	3.43			± 0.30
PIC RANGE		8			± 4	7			± 4	7			± 4

Ref: TEMP-0382 Rev.31 BACK / VERSO 9930083 -A

* For research use only in Canada and USA