

ABX Minotrol Retic



LOT RX 074

CONTROL

(Exp.) 2014-09-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual													
RETIC	%	1.3	1.3	1.3	± 0.7	5.8	5.8	5.8	± 2.3	11.0	11.0	11.0	± 3.3
RETIC	%	2.31	2.36	2.36	± 1.18	5.67	5.66	5.66	± 1.42	11.30	11.40	11.40	± 2.30
RETIC	10 ⁶ /µl	0.078	0.080	0.080	± 0.040	0.185	0.185	0.185	± 0.046	0.375	0.380	0.380	± 0.076
MFI MRI *		28.5	30.2	30.2	± 8.0	40.9	42.9	42.9	± 10.0	44.1	45.2	45.2	± 12.0
GR RBC	10 ⁶ /µl	3.37	3.37	3.37	± 0.30	3.27	3.27	3.27	± 0.30	3.32	3.33	3.33	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

ABX Minotrol Retic



LOT RX 074
Rev 1

CONTROL

(Exp.) 2014-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	5.8			± 2.3	11.0			± 3.3
RETIC	%	2.28			± 1.14	5.59			± 1.40	11.30			± 2.30
RETIC	10 ⁶ /µl	0.077			± 0.039	0.185			± 0.046	0.375			± 0.075
MFI MRI *		29.3			± 8.0	42.3			± 10.0	44.9			± 12.0
GR RBC	10 ⁶ /µl	3.38			± 0.30	3.31			± 0.30	3.32			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083 -A

* For research use only in Canada and USA