

ABX Minotrol Retic



LOT RX 114

CONTROL

(Exp.) 2015-01-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual													
RETIC	%	1.4	1.4	1.4	± 0.7	6.3	6.3	6.3	± 2.5	11.2	11.2	11.2	± 3.4
RETIC	%	2.25	2.45	2.45	± 1.23	5.54	5.71	5.71	± 1.43	11.40	11.50	11.50	± 2.30
RETIC	10 ⁶ /µl	0.076	0.083	0.083	± 0.042	0.183	0.190	0.190	± 0.048	0.382	0.390	0.390	± 0.078
MFI MRI *		27.2	28.8	28.8	± 8.0	37.2	39.5	39.5	± 10.0	44.3	46.4	46.4	± 12.0
GR RBC	10 ⁶ /µl	3.38	3.39	3.39	± 0.30	3.30	3.33	3.33	± 0.30	3.35	3.39	3.39	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

ABX Minotrol Retic



LOT RX 114
Rev 1

CONTROL

(Exp.) 2015-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	6.3			± 2.5	11.2			± 3.4
RETIC	%	2.38			± 1.19	5.71			± 1.43	11.50			± 2.30
RETIC	10 ⁶ /µl	0.081			± 0.041	0.188			± 0.047	0.384			± 0.077
MFI MRI *		26.5			± 8.0	38.0			± 10.0	44.8			± 12.0
GR RBC	10 ⁶ /µl	3.40			± 0.30	3.30			± 0.30	3.34			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083 -A

* For research use only in Canada and USA