


ABX Minotrol Retic

LOT		RX 035				CONTROL				<div><div></div><div>(Exp.) 2015-05-05</div><div>(YYYY - MM - DD)</div></div>						
		Rev 1		ABX												
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			1	TOLERANCES TOLERANCE	CONTROL			2	TOLERANCES TOLERANCE	CONTROL			3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA							
		120 RETIC	DX120	DX Nexus	120 RETIC		DX120	DX Nexus	120 RETIC	DX120		DX Nexus				
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	6.2	6.2	6.2	± 2.5	11.7	11.7	11.7	± 3.5			
RETIC	%	1.89	2.07	2.07	± 1.04	5.31	5.55	5.55	± 1.39	11.10	11.40	11.40	± 2.30			
RETIC	10 ⁶ /µl	0.064	0.070	0.070	± 0.035	0.177	0.186	0.186	± 0.047	0.376	0.388	0.388	± 0.077			
MFI MRI *		28.0	27.1	27.1	± 8.0	44.4	45.3	45.3	± 10.0	48.4	50.5	50.5	± 12.0			
GR RBC	10 ⁶ /µl	3.36	3.38	3.38	± 0.30	3.34	3.35	3.35	± 0.30	3.39	3.40	3.40	± 0.30			
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0			

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

ABX Minotrol Retic

LOT		RX 035		CONTROL		<div><div></div><div>(Exp.) 2015-05-05</div><div>(YYYY - MM - DD)</div></div>										
Rev 1																
PARAMETRES PARAMETERS	UNITES UNITS	ABX														
		CONTROL			1	TOLERANCES TOLERANCE	CONTROL			2	TOLERANCES TOLERANCE	CONTROL			3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA							
		XLR			XLR				XLR							
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	6.2			± 2.5	11.7			± 3.5			
RETIC	%	1.90			± 1.04	5.41			± 1.39	11.20			± 2.30			
RETIC	10 ⁶ /µl	0.064			± 0.035	0.181			± 0.047	0.380			± 0.077			
MFI MRI *		28.2			± 8.0	43.9			± 10.0	49.1			± 12.0			
GR RBC	10 ⁶ /µl	3.37			± 0.30	3.34			± 0.30	3.39			± 0.30			
PIC RANGE		8			± 4	8			± 4	8			± 4			

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A