

# ABX Minotrol Retic



**LOT** RX 055  
Rev 1

**CONTROL**

(Exp.) 2015-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.6	1.6	1.6	± 0.8	6.6	6.6	6.6	± 2.6	11.7	11.7	11.7	± 3.5
RETIC	%	1.99	2.02	2.02	± 1.01	5.40	5.71	5.71	± 1.43	11.10	11.50	11.50	± 2.30
RETIC	10 <sup>6</sup> /µl	0.067	0.068	0.068	± 0.034	0.183	0.194	0.194	± 0.049	0.372	0.389	0.389	± 0.078
MFI MRI *		31.9	32.0	32.0	± 8.0	43.7	45.4	45.4	± 10.0	48.9	49.5	49.5	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.37	3.37	3.37	± 0.30	3.38	3.39	3.39	± 0.30	3.35	3.38	3.38	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 055  
Rev 1

**CONTROL**

(Exp.) 2015-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.6</b>			± 0.8	<b>6.6</b>			± 2.6	<b>11.7</b>			± 3.5
RETIC	%	<b>2.01</b>			± 1.01	<b>5.53</b>			± 1.43	<b>11.20</b>			± 2.30
RETIC	10 <sup>6</sup> /µl	<b>0.068</b>			± 0.034	<b>0.187</b>			± 0.049	<b>0.377</b>			± 0.078
MFI MRI *		<b>29.6</b>			± 8.0	<b>44.4</b>			± 10.0	<b>49.4</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.37</b>			± 0.30	<b>3.39</b>			± 0.30	<b>3.37</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

\* For research use only in Canada and USA