

ABX Minotrol Retic



LOT RX 095
Rev 1

CONTROL

(Exp.) 2015-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	5.7	5.7	5.7	± 2.3	10.3	10.3	10.3	± 3.1
RETIC	%	2.43	2.44	2.44	± 1.22	5.65	5.87	5.87	± 1.47	11.70	11.80	11.80	± 2.40
RETIC	10 ⁶ /µl	0.083	0.084	0.084	± 0.042	0.194	0.203	0.203	± 0.051	0.401	0.405	0.405	± 0.081
MFI MRI *		33.5	33.5	33.5	± 8.0	45.1	44.5	44.5	± 10.0	49.9	48.9	48.9	± 12.0
GR RBC	10 ⁶ /µl	3.42	3.45	3.45	± 0.30	3.44	3.45	3.45	± 0.30	3.43	3.43	3.43	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 095
Rev 1

CONTROL

(Exp.) 2015-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	5.7			± 2.3	10.3			± 3.1
RETIC	%	2.44			± 1.22	5.67			± 1.42	11.50			± 2.30
RETIC	10 ⁶ /µl	0.083			± 0.042	0.193			± 0.048	0.390			± 0.078
MFI MRI *		35.9			± 8.0	46.4			± 10.0	49.8			± 12.0
GR RBC	10 ⁶ /µl	3.41			± 0.30	3.40			± 0.30	3.39			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA