

# ABX Minotrol Retic



**LOT** RX 430  
Rev 1

**CONTROL**

(Exp.) 2021-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%		1.4	1.4	± 0.7		6.1	6.1	± 2.4		11.1	11.1	± 3.3
RETIC	%		1.66	1.66	± 0.83		5.38	5.38	± 1.35		11.10	11.10	± 2.20
RETIC	10 <sup>6</sup> /µl		0.057	0.057	± 0.029		0.184	0.184	± 0.046		0.379	0.379	± 0.076
MFI MRI *	%		31.4	31.4	± 8.0		46.0	46.0	± 10.0		50.6	50.6	± 12.0
GR RBC	10 <sup>6</sup> /µl		3.46	3.46	± 0.30		3.42	3.42	± 0.30		3.41	3.41	± 0.30
PIC RANGE	-		27.0	27.0	± 5.0		27.0	27.0	± 5.0		27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

# ABX Minotrol Retic



**LOT** RX 430  
Rev 1

**CONTROL**

(Exp.) 2021-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>6.1</b>			± 2.4	<b>11.1</b>			± 3.3
RETIC	%	<b>1.64</b>			± 0.82	<b>5.55</b>			± 1.39	<b>11.56</b>			± 2.31
RETIC	10 <sup>6</sup> /µl	<b>0.054</b>			± 0.027	<b>0.181</b>			± 0.045	<b>0.378</b>			± 0.076
MFI MRI *	%	<b>32.0</b>			± 8.0	<b>45.2</b>			± 10.0	<b>48.7</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.31</b>			± 0.30	<b>3.26</b>			± 0.30	<b>3.27</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA