

ABX Minotrol Retic



LOT RX 413
Rev 1

CONTROL

(Exp.) 2018-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 0.7	4.8	4.8	4.8	± 2.0	9.8	9.8	9.8	± 3.0
RETIC	%	1.82	1.86	1.86	± 0.93	5.11	5.23	5.23	± 1.31	10.90	11.40	11.40	± 2.30
RETIC	10 ⁶ /µl	0.064	0.063	0.063	± 0.032	0.182	0.179	0.179	± 0.046	0.379	0.380	0.380	± 0.076
MFI MRI *		29.9	27.3	27.3	± 8.0	48.1	45.4	45.4	± 10.0	51.0	49.5	49.5	± 12.0
GR RBC	10 ⁶ /µl	3.54	3.41	3.41	± 0.30	3.57	3.43	3.43	± 0.30	3.48	3.33	3.33	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 413
Rev 1

CONTROL

(Exp.) 2018-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	4.8			± 2.0	9.8			± 3.0
RETIC	%	1.78			± 0.89	5.04			± 1.26	11.00			± 2.20
RETIC	10 ⁶ /µl	0.061			± 0.031	0.174			± 0.044	0.366			± 0.073
MFI MRI *		27.7			± 8.0	45.9			± 10.0	51.3			± 12.0
GR RBC	10 ⁶ /µl	3.42			± 0.30	3.45			± 0.30	3.33			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA