

ABX Minotrol Retic



LOT RX 432
Rev 1

CONTROL

(Exp.) 2022-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	± 0.7	5.5	5.5	± 2.2	12.2	12.2	± 3.7			
RETIC	%	1.55	1.55	± 0.78	4.79	4.79	± 1.20	9.94	9.94	± 1.99			
RETIC	10 ⁶ /µl	0.053	0.053	± 0.027	0.164	0.164	± 0.041	0.339	0.339	± 0.068			
MFI MRI *	%	30.7	30.7	± 8.0	43.6	43.6	± 10.0	48.5	48.5	± 12.0			
GR RBC	10 ⁶ /µl	3.43	3.43	± 0.30	3.42	3.42	± 0.30	3.41	3.41	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.40 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 432
Rev 1

CONTROL

(Exp.) 2022-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	5.5			± 2.2	12.2			± 3.7
RETIC	%	1.66			± 0.83	4.81			± 1.20	10.20			± 2.04
RETIC	10 ⁶ /µl	0.056			± 0.028	0.162			± 0.041	0.344			± 0.069
MFI MRI *	%	30.5			± 8.0	39.0			± 10.0	42.3			± 12.0
GR RBC	10 ⁶ /µl	3.38			± 0.30	3.37			± 0.30	3.37			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.40 BACK / VERSO 1300090305

* For research use only in Canada and USA