

ABX Minotrol Retic



LOT RX 422
Rev 2

CONTROL

(Exp.) 2020-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 0.7	6.0	6.0	6.0	± 2.4	11.5	11.5	11.5	± 3.5
RETIC	%	1.47	1.77	1.77	± 0.89	5.04	5.33	5.33	± 1.33	10.30	10.90	10.90	± 2.20
RETIC	10 ⁶ /µl	0.051	0.062	0.062	± 0.031	0.176	0.187	0.187	± 0.047	0.358	0.378	0.378	± 0.076
MFI MRI *	%	32.0	32.0	32.0	± 8.0	46.4	46.5	46.5	± 10.0	51.4	51.1	51.1	± 12.0
GR RBC	10 ⁶ /µl	3.49	3.51	3.51	± 0.30	3.49	3.50	3.50	± 0.30	3.48	3.47	3.47	± 0.30
PIC RANGE	-	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.37 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 422
Rev 2

CONTROL

(Exp.) 2020-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	6.0			± 2.4	11.5			± 3.5
RETIC	%	1.75			± 0.88	5.29			± 1.32	10.61			± 2.12
RETIC	10 ⁶ /µl	0.060			± 0.030	0.183			± 0.046	0.364			± 0.073
MFI MRI *	%	29.6			± 8.0	44.5			± 10.0	47.7			± 12.0
GR RBC	10 ⁶ /µl	3.45			± 0.30	3.45			± 0.30	3.43			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.37 BACK / VERSO 1219930083-A

* For research use only in Canada and USA