

ABX Minotrol Retic



LOT RX 420
Rev 1

CONTROL

(Exp.) 2020-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	5.3	5.3	5.3	± 2.1	11.3	11.3	11.3	± 3.4
RETIC	%	1.80	2.00	2.00	± 1.00	5.19	5.59	5.59	± 1.40	10.40	11.10	11.10	± 2.30
RETIC	10 ⁶ /µl	0.062	0.069	0.069	± 0.035	0.180	0.195	0.195	± 0.049	0.357	0.382	0.382	± 0.076
MFI MRI *		34.1	34.4	34.4	± 8.0	43.9	44.6	44.6	± 10.0	47.2	48.8	48.8	± 12.0
GR RBC	10 ⁶ /µl	3.44	3.45	3.45	± 0.30	3.46	3.48	3.48	± 0.30	3.43	3.44	3.44	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 420
Rev 1

CONTROL

(Exp.) 2020-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	5.3			± 2.1	11.3			± 3.4
RETIC	%	1.96			± 0.98	5.49			± 1.37	10.90			± 2.18
RETIC	10 ⁶ /µl	0.068			± 0.034	0.191			± 0.048	0.374			± 0.075
MFI MRI *		35.4			± 8.0	45.7			± 10.0	48.7			± 12.0
GR RBC	10 ⁶ /µl	3.45			± 0.30	3.48			± 0.30	3.43			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA