

ABX Minotrol Retic



LOT RX 400
Rev 1

CONTROL

(Exp.) 2016-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	6.9	6.9	6.9	± 2.8	11.3	11.3	11.3	± 3.4
RETIC	%	2.12	2.18	2.18	± 1.09	5.57	5.89	5.89	± 1.47	11.20	11.60	11.60	± 2.30
RETIC	10 ⁶ /µl	0.072	0.074	0.074	± 0.037	0.187	0.199	0.199	± 0.050	0.380	0.394	0.394	± 0.079
MFI MRI *		30.3	30.1	30.1	± 8.0	46.5	45.0	45.0	± 10.0	49.8	50.3	50.3	± 12.0
GR RBC	10 ⁶ /µl	3.39	3.41	3.41	± 0.30	3.36	3.38	3.38	± 0.30	3.39	3.40	3.40	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

ABX Minotrol Retic



LOT RX 400
Rev 1

CONTROL

(Exp.) 2016-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	6.9			± 2.8	11.3			± 3.4
RETIC	%	2.16			± 1.08	5.72			± 1.43	11.40			± 2.30
RETIC	10 ⁶ /µl	0.073			± 0.037	0.191			± 0.048	0.383			± 0.077
MFI MRI *		28.6			± 8.0	43.5			± 10.0	49.4			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.34			± 0.30	3.36			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA