


ABX Minotrol Retic



LOT	RX 408	CONTROL								(Exp.) 2018-01-05				
	Rev 2													(YYYY - MM - DD)
PARAMETRES PARAMETERS	UNITES UNITS	ABX												
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE	
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		
		120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus				
Méthode Manuelle / Manual														
RETIC	%	1.4	1.4	1.4	± 0.7	6.3	6.3	6.3	± 2.5	10.9	10.9	10.9	± 3.3	
RETIC	%	2.52	2.60	2.60	± 1.30	5.71	5.75	5.75	± 1.44	11.50	11.80	11.80	± 2.40	
RETIC	10 ⁶ /µl	0.087	0.091	0.091	± 0.046	0.198	0.200	0.200	± 0.050	0.398	0.408	0.408	± 0.082	
MFI MRI *		36.9	36.9	36.9	± 8.0	43.9	43.9	43.9	± 10.0	51.9	51.9	51.9	± 12.0	
GR RBC	10 ⁶ /µl	3.45	3.49	3.49	± 0.30	3.47	3.47	3.47	± 0.30	3.46	3.46	3.46	± 0.30	
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 408		CONTROL			(Exp.) 2018-01-05 (YYYY - MM - DD)								
Rev 2		ABX											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual													
RETIC	%	1.4			± 0.7	6.3			± 2.5	10.9			± 3.3
RETIC	%	2.24			± 1.12	5.50			± 1.38	11.60			± 2.32
RETIC	10 ⁶ /µl	0.077			± 0.039	0.189			± 0.047	0.396			± 0.079
MFI MRI *		35.1			± 8.0	41.1			± 10.0	45.8			± 12.0
GR RBC	10 ⁶ /µl	3.45			± 0.30	3.44			± 0.30	3.41			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A