

ABX Minotrol Retic



LOT RX 429
Rev 1

CONTROL

(Exp.) 2021-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%		1.4	1.4	± 1.0		5.5	5.5	± 2.2		10.9	10.9	± 3.3
RETIC	%		1.60	1.60	± 0.80		4.60	4.60	± 1.15		10.90	10.90	± 2.20
RETIC	10 ⁶ /µl		0.056	0.056	± 0.028		0.158	0.158	± 0.040		0.378	0.378	± 0.076
MFI MRI *	%		32.5	32.5	± 8.0		47.9	47.9	± 10.0		52.4	52.4	± 12.0
GR RBC	10 ⁶ /µl		3.50	3.50	± 0.30		3.43	3.43	± 0.30		3.47	3.47	± 0.30
PIC RANGE	-		27.0	27.0	± 5.0		27.0	27.0	± 5.0		27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 429
Rev 1

CONTROL

(Exp.) 2021-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 1.0	5.5			± 2.2	10.9			± 3.3
RETIC	%	1.76			± 0.88	5.09			± 1.28	10.62			± 2.13
RETIC	10 ⁶ /µl	0.060			± 0.030	0.171			± 0.043	0.360			± 0.072
MFI MRI *	%	29.7			± 8.0	41.4			± 10.0	46.9			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.35			± 0.30	3.39			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

* For research use only in Canada and USA