

ABX Minotrol Retic



LOT RX 406
Rev 1

CONTROL

(Exp.) 2017-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 1.0	5.7	5.7	5.7	± 2.3	11.6	11.6	11.6	± 3.5
RETIC	%	1.96	2.18	2.18	± 1.09	5.54	5.74	5.74	± 1.44	11.70	12.00	12.00	± 2.40
RETIC	10 ⁶ /µl	0.069	0.077	0.077	± 0.039	0.196	0.204	0.204	± 0.051	0.415	0.426	0.426	± 0.085
MFI MRI *		26.7	28.4	28.4	± 8.0	44.6	46.5	46.5	± 10.0	49.1	51.9	51.9	± 12.0
GR RBC	10 ⁶ /µl	3.54	3.53	3.53	± 0.30	3.54	3.56	3.56	± 0.30	3.55	3.55	3.55	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 406

CONTROL

(Exp.) 2017-09-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.5			± 1.0	5.7			± 2.3	11.6			± 3.5
RETIC	%	1.95			± 0.98	5.60			± 1.40	11.80			± 2.36
RETIC	10 ⁶ /µl	0.069			± 0.035	0.195			± 0.049	0.413			± 0.083
MFI MRI *		28.2			± 8.0	45.4			± 10.0	50.8			± 12.0
GR RBC	10 ⁶ /µl	3.52			± 0.30	3.49			± 0.30	3.50			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A