

# ABX Minotrol Retic



**LOT** RX 414  
Rev 1

**CONTROL**

(Exp.) 2019-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.8	5.3	5.3	5.3	± 2.6	10.6	10.6	10.6	± 3.2
RETIC	%	1.65	1.69	1.69	± 0.85	5.03	5.26	5.26	± 1.32	10.50	11.20	11.20	± 2.30
RETIC	10 <sup>6</sup> /µl	0.060	0.058	0.058	± 0.030	0.184	0.184	0.184	± 0.046	0.383	0.389	0.389	± 0.080
MFI MRI *		25.6	22.7	22.7	± 8.0	45.3	40.9	40.9	± 10.0	52.4	46.2	46.2	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.64	3.46	3.46	± 0.30	3.66	3.49	3.49	± 0.30	3.65	3.47	3.47	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

# ABX Minotrol Retic



**LOT** RX 414  
Rev 1

**CONTROL**

(Exp.) 2019-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.8	<b>5.3</b>			± 2.6	<b>10.6</b>			± 3.2
RETIC	%	<b>1.57</b>			± 0.79	<b>5.08</b>			± 1.27	<b>10.80</b>			± 2.16
RETIC	10 <sup>6</sup> /µl	<b>0.054</b>			± 0.027	<b>0.176</b>			± 0.044	<b>0.372</b>			± 0.074
MFI MRI *		<b>24.2</b>			± 8.0	<b>42.2</b>			± 10.0	<b>47.9</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.44</b>			± 0.30	<b>3.46</b>			± 0.30	<b>3.44</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA