

ABX Minotrol Retic



LOT RX 415
Rev 3

CONTROL

(Exp.) 2019-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 0.7	4.9	4.9	4.9	± 2.0	10.0	10.0	10.0	± 3.0
RETIC	%	1.81	1.82	1.82	± 0.91	5.07	5.24	5.24	± 1.31	10.30	10.90	10.90	± 2.20
RETIC	10 ⁶ /µl	0.064	0.063	0.063	± 0.032	0.183	0.183	0.183	± 0.046	0.370	0.379	0.379	± 0.076
MFI MRI *		27.5	26.4	26.4	± 8.0	46.7	46.7	46.7	± 10.0	51.6	51.6	51.6	± 12.0
GR RBC	10 ⁶ /µl	3.56	3.46	3.46	± 0.30	3.60	3.49	3.49	± 0.30	3.59	3.48	3.48	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 415
Rev 3

CONTROL

(Exp.) 2019-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
XLR			XLR			XLR			XLR				
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	4.9			± 2.0	10.0			± 3.0
RETIC	%	2.06			± 1.03	4.96			± 1.24	10.70			± 2.14
RETIC	10 ⁶ /µl	0.071			± 0.036	0.173			± 0.043	0.371			± 0.074
MFI MRI *		25.3			± 8.0	42.6			± 10.0	48.5			± 12.0
GR RBC	10 ⁶ /µl	3.46			± 0.30	3.48			± 0.30	3.47			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA