

ABX Minotrol Retic



LOT RX 409
Rev 1

CONTROL

(Exp.) 2018-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.9	5.5	5.5	5.5	± 2.2	11.4	11.4	11.4	± 3.4
RETIC	%	2.18	2.17	2.17	± 1.09	5.35	5.65	5.65	± 1.42	10.90	11.60	11.60	± 2.40
RETIC	10 ⁶ /µl	0.076	0.075	0.075	± 0.038	0.190	0.199	0.199	± 0.050	0.390	0.412	0.412	± 0.082
MFI MRI *		29.1	26.8	26.8	± 8.0	45.4	43.8	43.8	± 10.0	51.2	51.5	51.5	± 12.0
GR RBC	10 ⁶ /µl	3.48	3.45	3.45	± 0.30	3.56	3.53	3.53	± 0.30	3.58	3.55	3.55	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 409
Rev 1

CONTROL

(Exp.) 2018-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.5			± 0.9	5.5			± 2.2	11.4			± 3.4
RETIC	%	1.92			± 0.96	5.29			± 1.32	11.20			± 2.30
RETIC	10 ⁶ /µl	0.066			± 0.033	0.186			± 0.047	0.396			± 0.079
MFI MRI *		27.3			± 8.0	45.1			± 10.0	52.5			± 12.0
GR RBC	10 ⁶ /µl	3.45			± 0.30	3.52			± 0.30	3.54			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA