

ABX Minotrol Retic



LOT RX 072
Rev 1

CONTROL

(Exp.)

2012-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.7	1.7	1.7	± 0.9	6.1	6.1	6.1	± 2.4	10.7	10.7	10.7	± 3.2
RETIC	%	1.62	1.64	1.64	± 0.82	4.93	4.97	4.97	± 1.24	10.10	10.10	10.10	± 2.10
RETIC	10 ⁵ /µl	0.057	0.058	0.058	± 0.029	0.173	0.175	0.175	± 0.044	0.351	0.356	0.356	± 0.074
MFI MRI *		30.9	30.2	30.2	± 8.0	41.5	41.9	41.9	± 10.0	44.6	44.8	44.8	± 10.0
GR RBC	10 ⁶ /µl	3.54	3.54	3.54	± 0.30	3.51	3.52	3.52	± 0.30	3.48	3.52	3.52	± 0.30
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5

9930059-A

Ref: TEMP-0382 Rev.08

* For research use only in Canada and USA