

# ABX Minotrol Retic



**LOT** RX 053  
Rev 1

**CONTROL**

(Exp.) 2013-07-05  
(YYYY-MM-DD)

		ABX														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				TOLERANCES TOLERANCE	CONTROL				TOLERANCES TOLERANCE	CONTROL				TOLERANCES TOLERANCE
		1			2			3								
		PENTRA	PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus				
Méthode Manuelle / Manual																
RETIC	%	1.5	1.5	1.5	± 0.8	5.9	5.9	5.9	± 2.4	11.3	11.3	11.3	± 3.4			
RETIC	%	2.20	2.05	2.05	± 1.10	5.11	5.30	5.30	± 1.33	10.10	10.60	10.60	± 2.20			
RETIC	10 <sup>6</sup> /µl	0.077	0.072	0.072	± 0.039	0.176	0.186	0.186	± 0.047	0.347	0.367	0.367	± 0.076			
MFI MRI *		35.0	36.0	36.0	± 8.0	44.9	44.9	44.9	± 10.0	46.0	47.3	47.3	± 10.0			
GR RBC	10 <sup>6</sup> /µl	3.48	3.52	3.52	± 0.30	3.45	3.50	3.50	± 0.30	3.44	3.46	3.46	± 0.30			
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5			

9930075-A  
Ref: TEMP-0382 Rev.30

\* For research use only in Canada and USA