

ABX Minotrol Retic



LOT RX 033
Rev 1

CONTROL



2013-05-05
(YYYY-MM-DD)

		ABX															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1				TOLERANCES TOLERANCE	CONTROL 2				TOLERANCES TOLERANCE	CONTROL 3				TOLERANCES TOLERANCE	
		PENTRA	PENTRA	PENTRA			PENTRA	PENTRA	PENTRA			PENTRA	PENTRA	PENTRA			
		120 RETIC	DX120	DX Nexus			120 RETIC	DX120	DX Nexus			120 RETIC	DX120	DX Nexus			
Méthode Manuelle / Manual																	
RETIC	%	1.3	1.3	1.3	± 0.7	5.7	5.7	5.7	± 2.3	10.7	10.7	10.7	± 3.2				
RETIC	%	1.63	1.52	1.52	± 0.82	4.92	5.08	5.08	± 1.27	10.10	10.40	10.40	± 2.10				
RETIC	10 ⁶ /µl	0.056	0.053	0.053	± 0.028	0.168	0.174	0.174	± 0.044	0.344	0.355	0.355	± 0.071				
MFI MRI *		28.6	30.6	30.6	± 8.0	43.9	43.6	43.6	± 10.0	46.4	47.5	47.5	± 10.0				
GR RBC	10 ⁶ /µl	3.42	3.46	3.46	± 0.30	3.42	3.42	3.42	± 0.30	3.41	3.41	3.41	± 0.30				
PIC RANGE		25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5	25.5	25.5	25.5	± 4.5				

9930075-A
Ref: TEMP-0382 Rev.30

* For research use only in Canada and USA