

# ABX Minotrol Retic



**LOT** RX 016  
Rev 1

**CONTROL**

(Exp.) 2016-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 0.7	6.7	6.7	6.7	± 2.7	11.5	11.5	11.5	± 3.5
RETIC	%	2.16	2.20	2.20	± 1.10	5.77	5.73	5.73	± 1.44	11.50	11.70	11.70	± 2.40
RETIC	10 <sup>6</sup> /µl	0.074	0.075	0.075	± 0.038	0.201	0.199	0.199	± 0.050	0.404	0.411	0.411	± 0.084
MFI MRI *		30.5	30.4	30.4	± 8.0	44.5	45.8	45.8	± 10.0	48.5	49.7	49.7	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.43	3.41	3.41	± 0.30	3.48	3.48	3.48	± 0.30	3.51	3.51	3.51	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 016  
Rev 1

**CONTROL**

(Exp.) 2016-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1		2		3							
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	<b>1.3</b>			± 0.7	<b>6.7</b>			± 2.7	<b>11.5</b>			± 3.5
RETIC	%	<b>2.17</b>			± 1.09	<b>5.74</b>			± 1.44	<b>11.50</b>			± 2.30
RETIC	10 <sup>6</sup> /µl	<b>0.074</b>			± 0.037	<b>0.199</b>			± 0.050	<b>0.403</b>			± 0.081
MFI MRI *		<b>29.3</b>			± 8.0	<b>43.2</b>			± 10.0	<b>49.3</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.43</b>			± 0.30	<b>3.47</b>			± 0.30	<b>3.50</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

\* For research use only in Canada and USA