

ABX Minotrol Retic



LOT RX 015
Rev 1

CONTROL

 (Exp.) 2015-03-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX | | | | | | | | | | | |
|---------------------------------------|---------------------|-----------|--------|----------|-------------------------|-----------|--------|----------|-------------------------|-----------|--------|----------|-------------------------|
| | | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE |
| | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | |
| | | 120 RETIC | DX120 | DX Nexus | | 120 RETIC | DX120 | DX Nexus | | 120 RETIC | DX120 | DX Nexus | |
| Méthode Manuelle / Manual RETIC | % | 1.3 | 1.3 | 1.3 | ± 0.7 | 6.2 | 6.2 | 6.2 | ± 2.5 | 10.9 | 10.9 | 10.9 | ± 3.3 |
| RETIC | % | 2.15 | 2.04 | 2.04 | ± 1.08 | 5.60 | 5.68 | 5.68 | ± 1.42 | 11.40 | 11.90 | 11.90 | ± 2.40 |
| RETIC | 10 ⁶ /µl | 0.074 | 0.071 | 0.071 | ± 0.037 | 0.192 | 0.195 | 0.195 | ± 0.049 | 0.398 | 0.418 | 0.418 | ± 0.084 |
| MFI MRI * | | 27.1 | 27.9 | 27.9 | ± 8.0 | 43.6 | 43.4 | 43.4 | ± 10.0 | 49.0 | 48.3 | 48.3 | ± 12.0 |
| GR RBC | 10 ⁶ /µl | 3.43 | 3.46 | 3.46 | ± 0.30 | 3.42 | 3.43 | 3.43 | ± 0.30 | 3.49 | 3.51 | 3.51 | ± 0.30 |
| PIC RANGE | | 27.0 | 27.0 | 27.0 | ± 5.0 | 27.0 | 27.0 | 27.0 | ± 5.0 | 27.0 | 27.0 | 27.0 | ± 5.0 |

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

ABX Minotrol Retic



LOT RX 015
Rev 1

CONTROL

(Exp.) 2015-03-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX | | | | | | | | | | | |
|---------------------------------------|---------------------|--------------|--------|--------|-------------------------|--------------|--------|--------|-------------------------|--------------|--------|--|-------------------------|
| | | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE |
| | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | | |
| | | XLR | | | | XLR | | | | XLR | | | |
| Méthode Manuelle / Manual RETIC | % | 1.3 | | | ± 0.7 | 6.2 | | | ± 2.5 | 10.9 | | | ± 3.3 |
| RETIC | % | 1.92 | | | ± 1.08 | 5.48 | | | ± 1.42 | 11.50 | | | ± 2.40 |
| RETIC | 10 ⁶ /µl | 0.067 | | | ± 0.037 | 0.190 | | | ± 0.049 | 0.404 | | | ± 0.084 |
| MFI MRI * | | 28.0 | | | ± 8.0 | 43.1 | | | ± 10.0 | 48.0 | | | ± 12.0 |
| GR RBC | 10 ⁶ /µl | 3.47 | | | ± 0.30 | 3.46 | | | ± 0.30 | 3.51 | | | ± 0.30 |
| PIC RANGE | | 8 | | | ± 4 | 8 | | | ± 4 | 8 | | | ± 4 |

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083 -A