


ABX Minotrol Retic



LOT RX 114
Rev 1

CONTROL

 (Exp.) 2015-01-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX | | | | | | | | | | | |
|---------------------------------------|---------------------|-----------|--------|----------|-------------------------|-----------|--------|----------|-------------------------|-----------|--------|----------|-------------------------|
| | | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE |
| | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | |
| | | 120 RETIC | DX120 | DX Nexus | | 120 RETIC | DX120 | DX Nexus | | 120 RETIC | DX120 | DX Nexus | |
| Méthode Manuelle / Manual RETIC | % | 1.4 | 1.4 | 1.4 | ± 0.7 | 6.3 | 6.3 | 6.3 | ± 2.5 | 11.2 | 11.2 | 11.2 | ± 3.4 |
| RETIC | % | 2.25 | 2.45 | 2.45 | ± 1.23 | 5.54 | 5.71 | 5.71 | ± 1.43 | 11.40 | 11.50 | 11.50 | ± 2.30 |
| RETIC | 10 ⁶ /µl | 0.076 | 0.083 | 0.083 | ± 0.042 | 0.183 | 0.190 | 0.190 | ± 0.048 | 0.382 | 0.390 | 0.390 | ± 0.078 |
| MFI MRI * | | 27.2 | 28.8 | 28.8 | ± 8.0 | 37.2 | 39.5 | 39.5 | ± 10.0 | 44.3 | 46.4 | 46.4 | ± 12.0 |
| GR RBC | 10 ⁶ /µl | 3.38 | 3.39 | 3.39 | ± 0.30 | 3.30 | 3.33 | 3.33 | ± 0.30 | 3.35 | 3.39 | 3.39 | ± 0.30 |
| PIC RANGE | | 27.0 | 27.0 | 27.0 | ± 5.0 | 27.0 | 27.0 | 27.0 | ± 5.0 | 27.0 | 27.0 | 27.0 | ± 5.0 |

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 114
Rev 1

CONTROL

(Exp.) 2015-01-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX | | | | | | | | | | | |
|---------------------------------------|---------------------|--------------|--------|--------|-------------------------|--------------|--------|--------|-------------------------|--------------|--------|--|-------------------------|
| | | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE |
| | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | | |
| | | XLR | | | | XLR | | | | XLR | | | |
| Méthode Manuelle / Manual RETIC | % | 1.4 | | | ± 0.7 | 6.3 | | | ± 2.5 | 11.2 | | | ± 3.4 |
| RETIC | % | 2.38 | | | ± 1.19 | 5.71 | | | ± 1.43 | 11.50 | | | ± 2.30 |
| RETIC | 10 ⁶ /µl | 0.081 | | | ± 0.041 | 0.188 | | | ± 0.047 | 0.384 | | | ± 0.077 |
| MFI MRI * | | 26.5 | | | ± 8.0 | 38.0 | | | ± 10.0 | 44.8 | | | ± 12.0 |
| GR RBC | 10 ⁶ /µl | 3.40 | | | ± 0.30 | 3.30 | | | ± 0.30 | 3.34 | | | ± 0.30 |
| PIC RANGE | | 8 | | | ± 4 | 8 | | | ± 4 | 8 | | | ± 4 |

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083 -A

* For research use only in Canada and USA