

# ABX Minotrol Retic



**LOT** RX 424  
Rev 1

**CONTROL**

(Exp.) 2020-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 1.0	4.9	4.9	4.9	± 2.4	10.7	10.7	10.7	± 3.2
RETIC	%	1.40	1.52	1.52	± 0.76	4.23	4.71	4.71	± 1.18	9.90	10.50	10.50	± 2.10
RETIC	10 <sup>6</sup> /µl	0.048	0.053	0.053	± 0.027	0.146	0.162	0.162	± 0.041	0.327	0.354	0.354	± 0.071
MFI MRI *	%	28.0	28.7	28.7	± 8.0	45.1	44.3	44.3	± 10.0	48.1	48.6	48.6	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.40	3.46	3.46	± 0.30	3.45	3.45	3.45	± 0.30	3.30	3.37	3.37	± 0.30
PIC RANGE	-	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.37 FRONT / RECTO 1219930083-A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 424  
Rev 1

**CONTROL**

(Exp.) 2020-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.3</b>			± 1.0	<b>4.9</b>			± 2.4	<b>10.7</b>			± 3.2
RETIC	%	<b>1.69</b>			± 0.85	<b>4.81</b>			± 1.20	<b>10.47</b>			± 2.09
RETIC	10 <sup>6</sup> /µl	<b>0.057</b>			± 0.029	<b>0.163</b>			± 0.041	<b>0.341</b>			± 0.068
MFI MRI *	%	<b>25.4</b>			± 8.0	<b>42.0</b>			± 10.0	<b>47.4</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.39</b>			± 0.30	<b>3.38</b>			± 0.30	<b>3.26</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.37 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA