

ABX Minotrol Retic



LOT RX 431
Rev 1

CONTROL

(Exp.) 2021-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA		PENTRA		PENTRA		PENTRA		PENTRA		PENTRA	
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	± 0.8	5.4	5.4	± 2.2	10.9	10.9	± 3.3			
RETIC	%	1.64	1.64	± 0.82	4.88	4.88	± 1.22	10.30	10.30	± 2.10			
RETIC	10 ⁶ /µl	0.057	0.057	± 0.029	0.169	0.169	± 0.042	0.357	0.357	± 0.071			
MFI MRI *	%	30.6	30.6	± 8.0	47.0	47.0	± 10.0	48.5	48.5	± 12.0			
GR RBC	10 ⁶ /µl	3.48	3.48	± 0.30	3.46	3.46	± 0.30	3.47	3.47	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.40 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 431
Rev 1

CONTROL

(Exp.) 2021-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	5.4			± 2.2	10.9			± 3.3
RETIC	%	1.58			± 0.79	5.00			± 1.25	10.50			± 2.10
RETIC	10 ⁶ /µl	0.054			± 0.027	0.169			± 0.042	0.356			± 0.071
MFI MRI *	%	31.3			± 8.0	46.4			± 10.0	47.7			± 12.0
GR RBC	10 ⁶ /µl	3.40			± 0.30	3.38			± 0.30	3.39			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.40 BACK / VERSO 1300090305

* For research use only in Canada and USA