

ABX Minotrol Retic



LOT RX 402
Rev 1

CONTROL

(Exp.) 2017-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.8	5.7	5.7	5.7	± 2.3	11.2	11.2	11.2	± 3.4
RETIC	%	2.20	2.09	2.09	± 1.10	5.15	5.26	5.26	± 1.32	10.80	10.70	10.70	± 2.20
RETIC	10 ⁶ /µl	0.076	0.072	0.072	± 0.038	0.176	0.179	0.179	± 0.045	0.364	0.360	0.360	± 0.073
MFI MRI *		23.0	24.4	24.4	± 8.0	39.7	41.0	41.0	± 10.0	43.9	44.4	44.4	± 12.0
GR RBC	10 ⁶ /µl	3.45	3.44	3.44	± 0.30	3.42	3.41	3.41	± 0.30	3.37	3.36	3.36	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.35 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 402
Rev 1

CONTROL

(Exp.) 2017-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.8	5.7			± 2.3	11.2			± 3.4
RETIC	%	1.92			± 0.96	5.18			± 1.30	10.60			± 2.10
RETIC	10 ⁶ /µl	0.066			± 0.033	0.175			± 0.044	0.353			± 0.071
MFI MRI *		24.8			± 8.0	41.0			± 10.0	44.5			± 12.0
GR RBC	10 ⁶ /µl	3.46			± 0.30	3.37			± 0.30	3.33			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.35 BACK / VERSO 9930083-A

* For research use only in Canada and USA