

# ABX Minotrol Retic



**LOT** RX 412  
Rev 1

**CONTROL**

(Exp.) 2018-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.7	1.7	1.7	± 0.9	5.5	5.5	5.5	± 2.2	10.2	10.2	10.2	± 3.1
RETIC	%	2.20	2.21	2.21	± 1.11	5.25	5.49	5.49	± 1.37	10.60	11.10	11.10	± 2.30
RETIC	10 <sup>6</sup> /µl	0.077	0.076	0.076	± 0.039	0.185	0.189	0.189	± 0.047	0.373	0.384	0.384	± 0.077
MFI MRI *		32.0	32.2	32.2	± 8.0	43.3	43.3	43.3	± 10.0	47.6	48.9	48.9	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.48	3.43	3.43	± 0.30	3.52	3.45	3.45	± 0.30	3.52	3.46	3.46	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 412  
Rev 1

**CONTROL**

(Exp.) 2018-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.7</b>			± 0.9	<b>5.5</b>			± 2.2	<b>10.2</b>			± 3.1
RETIC	%	<b>2.07</b>			± 1.04	<b>5.35</b>			± 1.34	<b>11.00</b>			± 2.20
RETIC	10 <sup>6</sup> /µl	<b>0.070</b>			± 0.035	<b>0.183</b>			± 0.046	<b>0.375</b>			± 0.075
MFI MRI *		<b>32.5</b>			± 8.0	<b>41.8</b>			± 10.0	<b>46.0</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.39</b>			± 0.30	<b>3.42</b>			± 0.30	<b>3.41</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A