

# ABX Minotrol Retic



**LOT** RX 403  
Rev 1

**CONTROL**

(Exp.) 2017-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	5.5	5.5	5.5	± 2.2	10.9	10.9	10.9	± 3.3
RETIC	%	2.06	2.09	2.09	± 1.05	5.62	5.68	5.68	± 1.42	11.40	11.50	11.50	± 2.30
RETIC	10 <sup>6</sup> /µl	0.073	0.073	0.073	± 0.037	0.197	0.198	0.198	± 0.050	0.394	0.397	0.397	± 0.079
MFI MRI *		26.5	27.1	27.1	± 8.0	45.3	46.9	46.9	± 10.0	50.8	51.2	51.2	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.52	3.49	3.49	± 0.30	3.50	3.48	3.48	± 0.30	3.46	3.45	3.45	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.35 FRONT / RECTO 9930083 -A

# ABX Minotrol Retic



**LOT** RX 403  
Rev 1

**CONTROL**

(Exp.) 2017-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.8	<b>5.5</b>			± 2.2	<b>10.9</b>			± 3.3
RETIC	%	<b>2.05</b>			± 1.03	<b>5.71</b>			± 1.43	<b>11.20</b>			± 2.24
RETIC	10 <sup>6</sup> /µl	<b>0.071</b>			± 0.036	<b>0.198</b>			± 0.050	<b>0.384</b>			± 0.077
MFI MRI *		<b>25.0</b>			± 8.0	<b>45.5</b>			± 10.0	<b>51.3</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.48</b>			± 0.30	<b>3.46</b>			± 0.30	<b>3.43</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.35 BACK / VERSO 9930083-A

\* For research use only in Canada and USA