

ABX Minotrol Retic



LOT RX 423
Rev 1

CONTROL

(Exp.) 2020-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.8	1.8	1.8	± 1.0	5.4	5.4	5.4	± 2.2	10.6	10.6	10.6	± 3.5
RETIC	%	1.52	1.63	1.63	± 0.82	5.01	5.11	5.11	± 1.28	10.60	10.70	10.70	± 2.20
RETIC	10 ⁶ /µl	0.053	0.058	0.058	± 0.029	0.175	0.182	0.182	± 0.046	0.373	0.383	0.383	± 0.079
MFI MRI *	%	30.0	29.4	29.4	± 8.0	45.2	45.2	45.2	± 10.0	48.7	49.9	49.9	± 12.0
GR RBC	10 ⁶ /µl	3.48	3.55	3.55	± 0.30	3.50	3.57	3.57	± 0.30	3.52	3.58	3.58	± 0.30
PIC RANGE	-	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.37 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 423
Rev 1

CONTROL

(Exp.) 2020-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.8			± 1.0	5.4			± 2.2	10.6			± 3.5
RETIC	%	1.75			± 0.88	5.02			± 1.26	10.50			± 2.10
RETIC	10 ⁶ /µl	0.059			± 0.030	0.171			± 0.043	0.358			± 0.072
MFI MRI *	%	26.8			± 8.0	41.4			± 10.0	45.9			± 12.0
GR RBC	10 ⁶ /µl	3.38			± 0.30	3.40			± 0.30	3.41			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.37 BACK / VERSO 1219930083-A

* For research use only in Canada and USA