

ABX Minotrol Retic



LOT RX 434
Rev 1

CONTROL

(Exp.) 2022-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	± 0.7	5.9	5.9	± 2.4	11.4	11.4	± 3.4			
RETIC	%	1.59	1.59	± 0.80	4.62	4.62	± 1.16	10.40	10.40	± 2.08			
RETIC	10 ⁶ /µl	0.056	0.056	± 0.028	0.161	0.161	± 0.040	0.365	0.365	± 0.073			
MFI MRI *	%	30.2	30.2	± 8.0	45.8	45.8	± 10.0	51.7	51.7	± 12.0			
GR RBC	10 ⁶ /µl	3.52	3.52	± 0.30	3.49	3.49	± 0.30	3.51	3.51	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.40 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 434
Rev 1

CONTROL

(Exp.) 2022-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	5.9			± 2.4	11.4			± 3.4
RETIC	%	1.66			± 0.83	5.13			± 1.28	11.11			± 2.22
RETIC	10 ⁶ /µl	0.058			± 0.029	0.180			± 0.045	0.391			± 0.078
MFI MRI *	%	32.0			± 8.0	46.5			± 10.0	50.1			± 12.0
GR RBC	10 ⁶ /µl	3.52			± 0.30	3.50			± 0.30	3.52			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.40 BACK / VERSO 1300090305

* For research use only in Canada and USA