

ABX Minotrol Retic



LOT RX 405
Rev 1

CONTROL

(Exp.) 2017-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	6.0	6.0	6.0	± 2.4	11.7	11.7	11.7	± 3.5
RETIC	%	1.97	2.10	2.10	± 1.05	5.62	5.66	5.66	± 1.42	12.00	11.90	11.90	± 2.40
RETIC	10 ⁶ /µl	0.068	0.072	0.072	± 0.036	0.194	0.194	0.194	± 0.049	0.410	0.403	0.403	± 0.082
MFI MRI *		25.9	27.7	27.7	± 8.0	43.5	45.7	45.7	± 10.0	47.9	49.7	49.7	± 12.0
GR RBC	10 ⁶ /µl	3.44	3.43	3.43	± 0.30	3.46	3.42	3.42	± 0.30	3.42	3.39	3.39	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 405

CONTROL

(Exp.) 2017-07-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1		2		3							
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	6.0			± 2.4	11.7			± 3.5
RETIC	%	2.01			± 1.01	5.55			± 1.39	11.80			± 2.40
RETIC	10 ⁶ /µl	0.070			± 0.035	0.192			± 0.048	0.405			± 0.081
MFI MRI *		27.3			± 8.0	45.4			± 10.0	50.4			± 12.0
GR RBC	10 ⁶ /µl	3.50			± 0.30	3.46			± 0.30	3.43			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA