

ABX Minotrol Retic



LOT RX 435
Rev 1

CONTROL

(Exp.) 2022-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	± 0.7	5.2	5.2	± 2.1	10.1	10.1	± 3.0			
RETIC	%	1.59	1.59	± 0.80	4.97	4.97	± 1.24	10.20	10.20	± 2.04			
RETIC	10 ⁶ /µl	0.055	0.055	± 0.028	0.170	0.170	± 0.043	0.346	0.346	± 0.069			
MFI MRI *	%	28.9	28.9	± 8.0	47.6	47.6	± 10.0	51.6	51.6	± 12.0			
GR RBC	10 ⁶ /µl	3.47	3.47	± 0.30	3.43	3.43	± 0.30	3.39	3.39	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.41 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 435
Rev 1

CONTROL

(Exp.) 2022-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	5.2			± 2.1	10.1			± 3.0
RETIC	%	1.58			± 0.79	5.20			± 1.30	10.76			± 2.15
RETIC	10 ⁶ /µl	0.053			± 0.027	0.173			± 0.043	0.353			± 0.071
MFI MRI *	%	28.1			± 8.0	43.4			± 10.0	46.5			± 12.0
GR RBC	10 ⁶ /µl	3.37			± 0.30	3.33			± 0.30	3.28			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.41 BACK / VERSO 1300090305

* For research use only in Canada and USA