

ABX Minotrol Retic



LOT RX 426
Rev 1

CONTROL

(Exp.) 2021-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.8	5.8	5.8	5.8	± 2.3	11.6	11.6	11.6	± 3.8
RETIC	%	1.55	1.83	1.83	± 0.92	4.72	5.28	5.28	± 1.32	9.59	10.60	10.60	± 2.10
RETIC	10 ⁶ /µl	0.053	0.063	0.063	± 0.032	0.160	0.182	0.182	± 0.046	0.329	0.368	0.368	± 0.074
MFI MRI *	%	36.7	34.0	34.0	± 8.0	46.7	46.6	46.6	± 10.0	47.6	48.7	48.7	± 12.0
GR RBC	10 ⁶ /µl	3.39	3.44	3.44	± 0.30	3.40	3.44	3.44	± 0.30	3.43	3.47	3.47	± 0.30
PIC RANGE	-	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 426
Rev 1

CONTROL

(Exp.) 2021-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.8	5.8			± 2.3	11.6			± 3.8
RETIC	%	1.92			± 0.96	5.15			± 1.29	10.23			± 2.05
RETIC	10 ⁶ /µl	0.066			± 0.033	0.177			± 0.044	0.355			± 0.071
MFI MRI *	%	31.5			± 8.0	44.0			± 10.0	45.4			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.44			± 0.30	3.47			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

* For research use only in Canada and USA