

# ABX Minotrol Retic



**LOT** RX 401  
Rev 1

**CONTROL**

(Exp.) 2016-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.8	1.8	1.8	± 1.0	6.1	6.1	6.1	± 2.4	11.0	11.0	11.0	± 3.3
RETIC	%	2.14	2.15	2.15	± 1.08	5.69	5.78	5.78	± 1.45	11.50	11.50	11.50	± 2.30
RETIC	10 <sup>6</sup> /µl	0.075	0.075	0.075	± 0.038	0.197	0.200	0.200	± 0.050	0.397	0.394	0.394	± 0.079
MFI MRI *		27.6	27.3	27.3	± 8.0	47.7	46.7	46.7	± 10.0	49.9	48.9	48.9	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.51	3.50	3.50	± 0.30	3.47	3.46	3.46	± 0.30	3.45	3.43	3.43	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 401  
Rev 1

**CONTROL**

(Exp.) 2016-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.8</b>			± 1.0	<b>6.1</b>			± 2.4	<b>11.0</b>			± 3.3
RETIC	%	<b>2.14</b>			± 1.07	<b>5.61</b>			± 1.40	<b>11.30</b>			± 2.26
RETIC	10 <sup>6</sup> /µl	<b>0.075</b>			± 0.038	<b>0.195</b>			± 0.049	<b>0.390</b>			± 0.078
MFI MRI *		<b>26.7</b>			± 8.0	<b>45.6</b>			± 10.0	<b>47.6</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.52</b>			± 0.30	<b>3.47</b>			± 0.30	<b>3.45</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

\* For research use only in Canada and USA