

# ABX Minotrol Retic



**LOT** RX 035  
Rev 1

**CONTROL**

(Exp.) 2015-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	6.2	6.2	6.2	± 2.5	11.7	11.7	11.7	± 3.5
RETIC	%	1.89	2.07	2.07	± 1.04	5.31	5.55	5.55	± 1.39	11.10	11.40	11.40	± 2.30
RETIC	10 <sup>6</sup> /µl	0.064	0.070	0.070	± 0.035	0.177	0.186	0.186	± 0.047	0.376	0.388	0.388	± 0.077
MFI MRI *		28.0	27.1	27.1	± 8.0	44.4	45.3	45.3	± 10.0	48.4	50.5	50.5	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.36	3.38	3.38	± 0.30	3.34	3.35	3.35	± 0.30	3.39	3.40	3.40	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

# ABX Minotrol Retic



**LOT** RX 035  
Rev 1

**CONTROL**

(Exp.) 2015-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.8	<b>6.2</b>			± 2.5	<b>11.7</b>			± 3.5
RETIC	%	<b>1.90</b>			± 1.04	<b>5.41</b>			± 1.39	<b>11.20</b>			± 2.30
RETIC	10 <sup>6</sup> /µl	<b>0.064</b>			± 0.035	<b>0.181</b>			± 0.047	<b>0.380</b>			± 0.077
MFI MRI *		<b>28.2</b>			± 8.0	<b>43.9</b>			± 10.0	<b>49.1</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.37</b>			± 0.30	<b>3.34</b>			± 0.30	<b>3.39</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

\* For research use only in Canada and USA