

ABX Minotrol Retic




LOT RX 034		CONTROL							(Exp.) 2014-05-05 (YYYY - MM - DD)				
Rev 2		ABX											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual													
RETIC	%	1.3	1.3	1.3	± 0.7	5.3	5.3	5.3	± 2.2	10.5	10.5	10.5	± 3.2
RETIC	%	2.00	2.20	2.20	± 1.10	5.07	5.30	5.30	± 1.23	11.10	11.30	11.30	± 2.20
RETIC	10 ⁶ /µl	0.068	0.075	0.075	± 0.037	0.174	0.181	0.181	± 0.045	0.391	0.379	0.379	± 0.074
MFI MRI *		25.6	24.2	24.2	± 8.0	41.1	39.7	39.7	± 10.0	47.9	47.4	47.4	± 12.0
GR RBC	10 ⁶ /µl	3.40	3.40	3.40	± 0.30	3.43	3.41	3.41	± 0.30	3.35	3.34	3.34	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.33 FRONT / RECTO 9930083 -A

ABX Minotrol Retic

LOT RX 034		CONTROL			 (Exp.) 2014-05-05 (YYYY - MM - DD)								
Rev 2		ABX											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual													
RETIC	%	1.3			± 0.7	5.3			± 2.2	10.5			± 3.2
RETIC	%	1.87			± 0.95	4.94			± 1.23	10.90			± 2.20
RETIC	10 ⁶ /µl	0.063			± 0.032	0.164			± 0.042	0.360			± 0.072
MFI MRI *		22.9			± 6.0	37.8			± 10.0	43.9			± 12.0
GR RBC	10 ⁶ /µl	3.35			± 0.30	3.33			± 0.30	3.31			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.33 BACK / VERSO 9930083 -A