

ABX Minotrol Retic



LOT RX 016
Rev 1

CONTROL

(Exp.) 2016-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 0.7	6.7	6.7	6.7	± 2.7	11.5	11.5	11.5	± 3.5
RETIC	%	2.16	2.20	2.20	± 1.10	5.77	5.73	5.73	± 1.44	11.50	11.70	11.70	± 2.40
RETIC	10 ⁶ /µl	0.074	0.075	0.075	± 0.038	0.201	0.199	0.199	± 0.050	0.404	0.411	0.411	± 0.084
MFI MRI *		30.5	30.4	30.4	± 8.0	44.5	45.8	45.8	± 10.0	48.5	49.7	49.7	± 12.0
GR RBC	10 ⁶ /µl	3.43	3.41	3.41	± 0.30	3.48	3.48	3.48	± 0.30	3.51	3.51	3.51	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 016
Rev 1

CONTROL

(Exp.) 2016-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	6.7			± 2.7	11.5			± 3.5
RETIC	%	2.17			± 1.09	5.74			± 1.44	11.50			± 2.30
RETIC	10 ⁶ /µl	0.074			± 0.037	0.199			± 0.050	0.403			± 0.081
MFI MRI *		29.3			± 8.0	43.2			± 10.0	49.3			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.47			± 0.30	3.50			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA