

# ABX Minotrol Retic



**LOT** RX 407  
Rev 1

**CONTROL**

(Exp.) 2017-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	5.9	5.9	5.9	± 2.4	11.2	11.2	11.2	± 3.4
RETIC	%	2.30	2.24	2.24	± 1.15	5.49	5.73	5.73	± 1.43	10.80	11.30	11.30	± 2.30
RETIC	10 <sup>6</sup> /µl	0.078	0.077	0.077	± 0.039	0.187	0.195	0.195	± 0.049	0.373	0.391	0.391	± 0.078
MFI MRI *		27.7	30.1	30.1	± 8.0	39.5	40.9	40.9	± 10.0	43.6	43.7	43.7	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.41	3.42	3.42	± 0.30	3.40	3.41	3.41	± 0.30	3.45	3.46	3.46	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 407  
Rev 1

**CONTROL**

(Exp.) 2017-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>5.9</b>			± 2.4	<b>11.2</b>			± 3.4
RETIC	%	<b>2.28</b>			± 1.14	<b>5.45</b>			± 1.36	<b>10.90</b>			± 2.20
RETIC	10 <sup>6</sup> /µl	<b>0.078</b>			± 0.039	<b>0.186</b>			± 0.047	<b>0.377</b>			± 0.075
MFI MRI *		<b>28.8</b>			± 8.0	<b>39.9</b>			± 10.0	<b>44.5</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.42</b>			± 0.30	<b>3.41</b>			± 0.30	<b>3.46</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA