

ABX Minotrol Retic



LOT RX 433
Rev 1

CONTROL

(Exp.) 2022-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.7	1.7	± 0.9	5.5	5.5	± 2.2	11.0	11.0	± 3.3			
RETIC	%	1.54	1.54	± 0.77	4.77	4.77	± 1.19	9.61	9.61	± 1.92			
RETIC	10 ⁶ /µl	0.052	0.052	± 0.026	0.162	0.162	± 0.041	0.330	0.330	± 0.066			
MFI MRI *	%	30.4	30.4	± 8.0	47.9	47.9	± 10.0	50.4	50.4	± 12.0			
GR RBC	10 ⁶ /µl	3.40	3.40	± 0.30	3.40	3.40	± 0.30	3.43	3.43	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.40 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 433
Rev 1

CONTROL

(Exp.) 2022-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.7			± 0.9	5.5			± 2.2	11.0			± 3.3
RETIC	%	1.55			± 0.78	5.03			± 1.26	10.20			± 2.04
RETIC	10 ⁶ /µl	0.052			± 0.026	0.170			± 0.043	0.347			± 0.069
MFI MRI *	%	29.9			± 8.0	42.2			± 10.0	45.8			± 12.0
GR RBC	10 ⁶ /µl	3.37			± 0.30	3.37			± 0.30	3.40			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.40 BACK / VERSO 1300090305

* For research use only in Canada and USA