

ABX Minotrol Retic



LOT RX 414
Rev 1

CONTROL

(Exp.) 2019-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.8	5.3	5.3	5.3	± 2.6	10.6	10.6	10.6	± 3.2
RETIC	%	1.65	1.69	1.69	± 0.85	5.03	5.26	5.26	± 1.32	10.50	11.20	11.20	± 2.30
RETIC	10 ⁶ /µl	0.060	0.058	0.058	± 0.030	0.184	0.184	0.184	± 0.046	0.383	0.389	0.389	± 0.080
MFI MRI *		25.6	22.7	22.7	± 8.0	45.3	40.9	40.9	± 10.0	52.4	46.2	46.2	± 12.0
GR RBC	10 ⁶ /µl	3.64	3.46	3.46	± 0.30	3.66	3.49	3.49	± 0.30	3.65	3.47	3.47	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 414
Rev 1

CONTROL

(Exp.) 2019-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.8	5.3			± 2.6	10.6			± 3.2
RETIC	%	1.57			± 0.79	5.08			± 1.27	10.80			± 2.16
RETIC	10 ⁶ /µl	0.054			± 0.027	0.176			± 0.044	0.372			± 0.074
MFI MRI *		24.2			± 8.0	42.2			± 10.0	47.9			± 12.0
GR RBC	10 ⁶ /µl	3.44			± 0.30	3.46			± 0.30	3.44			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA