

ABX Minotrol Retic



LOT RX 405
Rev 1

CONTROL

(Exp.) 2017-07-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX | | | | | | | | | | | |
|---------------------------------------|---------------------|----------|-----------|--------|-------------------------|-----------|--------|----------|-------------------------|---------|--------|--------|-------------------------|
| | | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE |
| | | 1 | | | | 2 | | | | 3 | | | |
| | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | |
| 120 RETIC | DX120 | DX Nexus | 120 RETIC | DX120 | DX Nexus | 120 RETIC | DX120 | DX Nexus | | | | | |
| Méthode Manuelle / Manual RETIC | % | 1.5 | 1.5 | 1.5 | ± 0.8 | 6.0 | 6.0 | 6.0 | ± 2.4 | 11.7 | 11.7 | 11.7 | ± 3.5 |
| RETIC | % | 1.97 | 2.10 | 2.10 | ± 1.05 | 5.62 | 5.66 | 5.66 | ± 1.42 | 12.00 | 11.90 | 11.90 | ± 2.40 |
| RETIC | 10 ⁶ /µl | 0.068 | 0.072 | 0.072 | ± 0.036 | 0.194 | 0.194 | 0.194 | ± 0.049 | 0.410 | 0.403 | 0.403 | ± 0.082 |
| MFI MRI * | | 25.9 | 27.7 | 27.7 | ± 8.0 | 43.5 | 45.7 | 45.7 | ± 10.0 | 47.9 | 49.7 | 49.7 | ± 12.0 |
| GR RBC | 10 ⁶ /µl | 3.44 | 3.43 | 3.43 | ± 0.30 | 3.46 | 3.42 | 3.42 | ± 0.30 | 3.42 | 3.39 | 3.39 | ± 0.30 |
| PIC RANGE | | 27.0 | 27.0 | 27.0 | ± 5.0 | 27.0 | 27.0 | 27.0 | ± 5.0 | 27.0 | 27.0 | 27.0 | ± 5.0 |

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 405

CONTROL

(Exp.) 2017-07-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX | | | | | | | | | | | |
|---------------------------------------|---------------------|--------------|--------|--------|-------------------------|--------------|--------|--------|-------------------------|--------------|--------|--|-------------------------|
| | | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE | CONTROL | | | TOLERANCES TOLERANCE |
| | | 1 | | 2 | | 3 | | | | | | | |
| | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | PENTRA | | PENTRA | PENTRA | | |
| XLR | | | XLR | | | XLR | | | | | | | |
| Méthode Manuelle / Manual RETIC | % | 1.5 | | | ± 0.8 | 6.0 | | | ± 2.4 | 11.7 | | | ± 3.5 |
| RETIC | % | 2.01 | | | ± 1.01 | 5.55 | | | ± 1.39 | 11.80 | | | ± 2.40 |
| RETIC | 10 ⁶ /µl | 0.070 | | | ± 0.035 | 0.192 | | | ± 0.048 | 0.405 | | | ± 0.081 |
| MFI MRI * | | 27.3 | | | ± 8.0 | 45.4 | | | ± 10.0 | 50.4 | | | ± 12.0 |
| GR RBC | 10 ⁶ /µl | 3.50 | | | ± 0.30 | 3.46 | | | ± 0.30 | 3.43 | | | ± 0.30 |
| PIC RANGE | | 8 | | | ± 4 | 8 | | | ± 4 | 8 | | | ± 4 |

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A