

# ABX Minotrol Retic



**LOT** RX 406  
Rev 1

**CONTROL**

(Exp.) 2017-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 1.0	5.7	5.7	5.7	± 2.3	11.6	11.6	11.6	± 3.5
RETIC	%	1.96	2.18	2.18	± 1.09	5.54	5.74	5.74	± 1.44	11.70	12.00	12.00	± 2.40
RETIC	10 <sup>6</sup> /µl	0.069	0.077	0.077	± 0.039	0.196	0.204	0.204	± 0.051	0.415	0.426	0.426	± 0.085
MFI MRI *		26.7	28.4	28.4	± 8.0	44.6	46.5	46.5	± 10.0	49.1	51.9	51.9	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.54	3.53	3.53	± 0.30	3.54	3.56	3.56	± 0.30	3.55	3.55	3.55	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

# ABX Minotrol Retic



**LOT** RX 406  
Rev 1

**CONTROL**

(Exp.) 2017-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 1.0	<b>5.7</b>			± 2.3	<b>11.6</b>			± 3.5
RETIC	%	<b>1.95</b>			± 0.98	<b>5.60</b>			± 1.40	<b>11.80</b>			± 2.36
RETIC	10 <sup>6</sup> /µl	<b>0.069</b>			± 0.035	<b>0.195</b>			± 0.049	<b>0.413</b>			± 0.083
MFI MRI *		<b>28.2</b>			± 8.0	<b>45.4</b>			± 10.0	<b>50.8</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.52</b>			± 0.30	<b>3.49</b>			± 0.30	<b>3.50</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA